



6 Step-by-step guide

on how to use the antigen self-test-kit

COVID-19 Antigen Test FLOWFLEX




1. Insert the swab **approx. 2,5 cm** into one nostril
 Rotate the swab **5 times**, gently pressing it against the mucous membrane
 Repeat the procedure in the other nostril

Unscrew the tube with buffer solution without squeezing it

Insert the swab into the tube and stir for **30 seconds**

Turn the swab at least **5 times** and squeeze the tube while doing so. Be careful not to spurt liquid out of the tube.




2.

Remove the stick while squeezing the tube

Screw the cap with dropper tip tightly on to the tube

Swirl and shake the tube




3.

Take the test cassette out of the packaging and place it flat on the table

Remove the top from the dropper tip

Carefully squeeze the tube and drop **4 drops** onto the test cassette [chamber S]

Important: The test cassette is sensitive to movement after the test liquid has been dripped onto it. In order not to falsify the test result, the cassette must therefore not be turned, shaken or, for example, carried from one room to another.



4.




5.

Wait for **15 minutes**
 Read the result

Interpret the result

6.



POSITIVE? → Call 1450!

NEGATIVE? Stay healthy!

Disclaimer: This guide is intended to provide a clear and easy-to-understand instruction on how to use the self-test kit to perform a rapid anterior-nasal COVID-19 antigen test. Please note that COVID-19 antigen rapid tests must always be carried out as specified in the instructions for use of the manufacturer or distributor. Please read the enclosed instructions for use carefully before using a COVID-19 antigen rapid test and follow the instructions. The Austrian Chamber of Pharmacists (Österreichische Apothekerkammer) assumes no liability for damage of any kind that may result from following the content of these instructions.

